



Certificate of Traceability

Manufacturer Certification

This certifies that **Boditech Med Inc.**, using **Immuno-assay on ichroma II** has participated in and successfully completed the NGSP certification for manufacturers and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **Reference Material Institute for Clinical Chemistry Standards (ReCCS) ASRL#1**

The system evaluated was:

Instrument: Ichroma II	Calibrator Lot: 609102
Reagent Lot: AAMKB27	Calibrator Assigned Values: 5.6%, 8.5%, 11.5%, 14%

Date of Certification: January 1, 2017

Certification Expires: January 1, 2018

NGSP Steering Committee Chair

NGSP Network Coordinator

SRL director/ supervisor